

Triad Group has initiated a voluntary product recall involving ALL LOTS of ALCOHOL PREP PADS, ALCOHOL SWABS, and ALCOHOL SWABSTICKS manufactured by TRIAD GROUP. This recall involves those products marked as STERILE as well as non-sterile products. This recall has been initiated due to concerns about potential contamination with Bacillus cereus.

They were distributed nationwide to retail pharmacies and are packaged in individual packets and sold in retail pharmacies in a box of 100 packets. The affected Alcohol Prep Pads, Alcohol Swabs and Alcohol Swabsticks can be identified by either Triad Group, listed as the manufacturer, or the products are manufactured for a third party and use the names listed below in their packaging: Walgreens, CVS. These products were distributed in the US.

DO NOT RETURN THE PRODUCT ON YOUR OWN. Call Triad Customer Service for a return authorization number at: 262-538-2900